
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000146
8:11 17/06/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:40
PRESSURIZING 3:43
STERILIZING 4:52
136.0 C 220kPa 4:52
136.4 C 224kPa 5:22
136.4 C 224kPa 5:52
136.4 C 224kPa 6:22
136.4 C 222kPa 6:52
136.2 C 222kPa 7:22
136.2 C 222kPa 7:52
136.4 C 221kPa 8:22
136.4 C 221kPa 8:22
VENTING 8:22
AIR DRYING 9:18
CYCLE COMPLETE 9:21

Digital Signature #
0018A7BC08EAA238
