
Midmark M11 - v1.0.0
Total Cycles: 5342

Name: RITTER 1
Sterilizer ID: V2178198
02/19/2025 11:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.5 F 0.0
2:00 138.4 F 1.4
4:00 141.8 F 4.0
6:00 160.4 F 7.2
8:00 182.6 F 11.0
10:00 213.4 F 2.9
12:00 223.9 F 6.0
14:00 228.6 F 10.1
16:00 235.3 F 8.8
18:00 239.7 F 10.9
20:00 236.1 F 8.5
22:00 249.1 F 14.5
24:00 259.6 F 20.3
26:00 269.5 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.4 F 29.1
2:00 271.5 F 28.4
3:00 272.4 F 29.2
4:00 272.2 F 28.7

Min 271.3 F 28.4
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.7 F 0.0
5:00 227.2 F 0.0
10:00 229.2 F 0.0
15:00 229.9 F 0.0
20:00 208.2 F 0.0
25:00 190.6 F 0.0
30:00 198.4 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 27:56
STERILIZING: 4:00
VENTING: 3:60
DRYING: 30:00
TOTAL CYCLE: 01:06:44
CYCLE COMPLETE

02/19/2025 12:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____