
Midmark M11 - v1.0.0
Total Cycles: 678

Name: MIDMARK
Sterilizer ID: V2178198
02/18/2021 03:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.5 F 0.0
2:00 132.5 F 1.0
4:00 158.7 F 3.6
6:00 185.5 F 8.0
8:00 208.3 F 13.4
10:00 219.9 F 6.9
12:00 232.8 F 12.1
14:00 237.1 F 9.0
16:00 237.9 F 9.2
18:00 248.3 F 14.0
20:00 260.3 F 21.0
22:00 271.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.1 F 28.6
2:00 273.0 F 29.1
3:00 272.9 F 29.3
4:00 271.7 F 28.6

Min 271.7 F 28.5
Max 273.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.2 F 0.0
5:00 241.1 F 0.0
10:00 223.2 F 0.0
15:00 232.2 F 0.0
20:00 220.0 F 0.0
25:00 236.2 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 22:33
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:00:45
CYCLE COMPLETE

02/18/2021 04:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____