

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 008365  
8:45 27/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:54  
PRESSURIZING 1:57  
STERILIZING 3:00  
136.0 C 220kPa 3:00  
136.0 C 221kPa 3:30  
136.4 C 223kPa 4:00  
136.4 C 220kPa 4:30  
136.4 C 220kPa 5:00  
136.3 C 221kPa 5:30  
136.4 C 221kPa 6:00  
136.3 C 221kPa 6:30  
136.3 C 221kPa 6:30  
VENTING 6:30  
AIR DRYING 7:18  
CYCLE COMPLETE 37:18

-----  
Digital Signature #  
C727AE59E068583C  
-----