

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002086  
14:02 23/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:36  
STERILIZING 3:39  
136.0 C 221kPa 3:39  
136.4 C 226kPa 4:09  
136.3 C 225kPa 4:39  
136.4 C 225kPa 5:09  
136.4 C 226kPa 5:39  
136.2 C 222kPa 6:09  
136.4 C 224kPa 6:39  
136.4 C 226kPa 7:09  
136.4 C 226kPa 7:09  
VENTING 7:09  
AIR DRYING 8:05  
CYCLE COMPLETE 8:20  
-----

Digital Signature #  
F95817B0B30386A8  
-----