
Midmark M11 - v1.0.5
Total Cycles: 688

Name: RITTER2
Sterilizer ID: V2313228
01/12/2022 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.9 F 0.0
2:00 168.8 F 3.5
4:00 188.1 F 7.9
6:00 209.7 F 13.4
8:00 226.3 F 8.6
10:00 235.4 F 10.6
12:00 243.2 F 12.7
14:00 242.7 F 12.0
16:00 256.1 F 19.1
18:00 268.0 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.0 F 29.4
2:00 271.5 F 29.5
3:00 272.0 F 29.2
4:00 271.7 F 29.7

Min 271.0 F 28.8
Max 272.3 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.9 F 0.0
5:00 227.2 F 0.0
10:00 229.1 F 0.0
15:00 227.1 F 0.0
20:00 226.8 F 0.0
25:00 231.5 F 0.0
30:00 231.2 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 19:13
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:51
CYCLE COMPLETE

01/12/2022 02:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____