

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5395

Name: RITTER 1  
Sterilizer ID: V2178198  
03/06/2025 03:33 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 86.4 F 0.0  
2:00 87.3 F 0.6  
4:00 112.4 F 2.5  
6:00 137.1 F 5.2  
8:00 157.4 F 8.2  
10:00 176.2 F 11.4  
12:00 197.4 F 3.6  
14:00 212.0 F 7.1  
16:00 219.7 F 10.7  
18:00 228.3 F 7.1  
20:00 236.2 F 10.7  
22:00 233.3 F 7.8  
24:00 245.9 F 13.8  
26:00 256.5 F 19.5  
28:00 266.5 F 25.7  
30:00 271.0 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.7 F 29.6  
1:00 272.4 F 29.7  
2:00 272.3 F 29.8  
3:00 272.2 F 30.0  
4:00 272.1 F 30.1  
  
Min 271.8 F 29.5  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.0 F 0.0  
5:00 201.6 F 0.0  
10:00 178.8 F 0.0  
15:00 168.1 F 0.0  
20:00 166.2 F 0.0  
25:00 164.0 F 0.0  
30:00 162.8 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 30:28  
STERILIZING: 4:00  
VENTING: 3:56  
DRYING: 30:00  
TOTAL CYCLE: 01:09:11  
CYCLE COMPLETE

03/06/2025 04:42 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_