
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005779
11:12 19/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:33
STERILIZING 3:37
136.0 C 220kPa 3:37
136.3 C 222kPa 4:07
136.4 C 225kPa 4:37
136.2 C 221kPa 5:07
136.4 C 225kPa 5:37
136.2 C 219kPa 6:07
136.4 C 225kPa 6:37
136.4 C 223kPa 7:07
136.4 C 223kPa 7:07
VENTING 7:07
AIR DRYING 8:00
CYCLE COMPLETE 8:06

Digital Signature #
DBA6C977D4E921B4
