
Midmark M11 - v1.0.5
Total Cycles: 2182

Name: RITTER2
Sterilizer ID: V2313228
03/27/2023 04:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.5 F 0.0
2:00 200.8 F 4.4
4:00 220.8 F 10.3
6:00 225.7 F 6.2
8:00 239.0 F 12.7
10:00 239.7 F 11.2
12:00 239.7 F 10.6
14:00 254.0 F 17.9
16:00 266.3 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 272.2 F 29.3
2:00 272.3 F 29.8
3:00 271.8 F 29.1
4:00 272.5 F 29.8

Min 271.5 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 233.4 F 0.0
10:00 235.3 F 0.0
15:00 230.6 F 0.0
20:00 227.6 F 0.0
25:00 229.3 F 0.0
30:00 230.2 F 0.0
Drying Complete

FILLING: 2:46
HEATING: 17:29
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:31
CYCLE COMPLETE

03/27/2023 05:50 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____