
Midmark M11 - v1.0.0
Total Cycles: 2043

Name: RITTER 1
Sterilizer ID: V2178198
06/22/2022 09:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.5 F 0.0
2:00 136.2 F 0.1
4:00 166.1 F 1.5
6:00 197.9 F 3.9
8:00 203.6 F 7.5
10:00 210.9 F 12.1
12:00 222.3 F 5.9
14:00 232.6 F 10.5
16:00 232.2 F 7.6
18:00 243.8 F 12.8
20:00 241.6 F 11.1
22:00 254.4 F 17.5
24:00 265.4 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.6 F 28.9
2:00 271.4 F 28.1
3:00 272.4 F 29.0
4:00 271.9 F 28.5

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.2 F 0.0
5:00 214.7 F 0.0
10:00 199.3 F 0.0
15:00 197.3 F 0.0
20:00 238.0 F 0.0
25:00 242.9 F 0.0
30:00 231.9 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 25:50
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:03:48
CYCLE COMPLETE

06/22/2022 10:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____