

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1969

Name: RITTER 1  
Sterilizer ID: V2178198  
05/30/2022 07:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.8 F 0.0  
2:00 78.8 F 0.1  
4:00 88.2 F 0.9  
6:00 115.5 F 3.1  
8:00 146.7 F 5.7  
10:00 169.9 F 8.7  
12:00 199.3 F 12.3  
14:00 206.8 F 3.6  
16:00 218.1 F 7.4  
18:00 229.0 F 11.2  
20:00 227.4 F 6.3  
22:00 239.3 F 11.3  
24:00 235.1 F 8.4  
26:00 248.2 F 14.5  
28:00 259.2 F 20.6  
30:00 269.1 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.4 F 29.3  
2:00 271.9 F 28.6  
3:00 271.5 F 28.8  
4:00 272.4 F 29.1  
  
Min 271.1 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.8 F 0.0  
5:00 200.5 F 0.0  
10:00 192.4 F 0.0  
15:00 194.4 F 0.0  
20:00 230.0 F 0.0  
25:00 230.0 F 0.0  
30:00 230.5 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 31:11  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:09:05  
CYCLE COMPLETE

05/30/2022 08:32 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_