
Midmark M11 - v1.0.5
Total Cycles: 1353

Name: RITTER2
Sterilizer ID: V2313228
07/27/2022 07:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.9 F 0.0
2:00 77.8 F 0.0
4:00 89.9 F 0.0
6:00 97.4 F 0.0
8:00 139.0 F 1.1
10:00 170.7 F 4.5
12:00 190.3 F 8.1
14:00 207.0 F 11.8
16:00 219.7 F 5.3
18:00 229.2 F 9.5
20:00 237.8 F 11.1
22:00 239.1 F 11.1
24:00 236.0 F 9.0
26:00 249.3 F 15.2
28:00 260.0 F 21.3
30:00 270.0 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.6 F 29.6
2:00 271.4 F 28.7
3:00 272.4 F 29.7
4:00 272.1 F 29.1

Min 271.1 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 225.5 F 0.0
10:00 193.7 F 0.0
15:00 195.3 F 0.0
20:00 217.8 F 0.0
25:00 222.8 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:59
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:08:52
CYCLE COMPLETE

07/27/2022 08:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____