
Midmark M11 - v1.0.0
Total Cycles: 3258

Name: RITTER 1
Sterilizer ID: V2178198
05/18/2023 02:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.0 F 0.0
2:00 138.7 F 0.0
4:00 151.5 F 1.1
6:00 180.4 F 5.2
8:00 203.7 F 9.5
10:00 228.9 F 10.9
12:00 235.7 F 9.7
14:00 234.4 F 8.8
16:00 244.4 F 12.3
18:00 240.7 F 10.4
20:00 253.5 F 16.8
22:00 264.6 F 23.4
24:00 270.1 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 271.7 F 28.9
2:00 271.7 F 28.8
3:00 271.9 F 28.8
4:00 272.1 F 28.9

Min 271.5 F 28.8
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.9 F 0.0
5:00 234.8 F 0.0
10:00 227.2 F 0.0
15:00 222.7 F 0.0
20:00 223.1 F 0.0
25:00 227.2 F 0.0
30:00 230.7 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:45
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:02:57
CYCLE COMPLETE

05/18/2023 03:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____