

-----  
Midmark M11 - v1.0.5  
Total Cycles: 37

Name: RITTER2  
Sterilizer ID: V2313228  
05/20/2021 10:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 159.1 F 0.0  
2:00 162.0 F 2.7  
4:00 173.7 F 6.3  
6:00 198.3 F 10.7  
8:00 217.5 F 5.2  
10:00 233.5 F 10.1  
12:00 235.2 F 8.9  
14:00 244.8 F 13.5  
16:00 244.7 F 12.8  
18:00 256.5 F 19.2  
20:00 267.2 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.8  
1:00 271.9 F 29.4  
2:00 273.0 F 29.8  
3:00 272.7 F 29.6  
4:00 272.8 F 29.3

Min 271.1 F 28.6  
Max 273.3 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.3 F 0.0  
5:00 204.9 F 0.0  
10:00 196.7 F 0.0  
15:00 195.2 F 0.0  
20:00 193.0 F 0.0  
25:00 188.8 F 0.0  
30:00 184.4 F 0.0  
Drying Complete

FILLING: 1:43  
HEATING: 21:36  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:00:38  
CYCLE COMPLETE

05/20/2021 11:36 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_