
Midmark M11 - v1.0.5
Total Cycles: 2477

Name: RITTER2
Sterilizer ID: V2313228
06/07/2023 09:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.8 F 0.0
2:00 138.4 F 0.0
4:00 146.1 F 0.7
6:00 174.3 F 4.8
8:00 206.5 F 9.4
10:00 217.8 F 8.8
12:00 226.0 F 9.2
14:00 234.9 F 10.5
16:00 239.7 F 11.6
18:00 238.1 F 10.0
20:00 251.2 F 16.6
22:00 262.6 F 23.2
24:00 272.8 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.6 F 30.0
2:00 271.6 F 29.2
3:00 272.3 F 30.0
4:00 272.4 F 29.8

Min 271.2 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 207.6 F 0.0
10:00 224.6 F 0.0
15:00 223.6 F 0.0
20:00 228.1 F 0.0
25:00 229.6 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 24:22
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:02:23
CYCLE COMPLETE

06/07/2023 10:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____