

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2857

Name: RITTER 1  
Sterilizer ID: V2178198  
02/06/2023 08:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.8 F 0.0  
2:00 76.6 F 0.0  
4:00 98.5 F 0.0  
6:00 116.4 F 0.2  
8:00 143.0 F 1.4  
10:00 173.6 F 4.5  
12:00 194.0 F 7.9  
14:00 210.0 F 11.8  
16:00 222.4 F 4.8  
18:00 232.0 F 9.1  
20:00 239.9 F 13.5  
22:00 237.9 F 10.3  
24:00 236.2 F 8.0  
26:00 248.9 F 14.5  
28:00 260.3 F 20.9  
30:00 270.8 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 271.8 F 28.4  
2:00 272.4 F 29.0  
3:00 271.2 F 28.4  
4:00 272.6 F 29.1  
  
Min 271.0 F 28.1  
Max 273.0 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.9 F 0.0  
5:00 203.4 F 0.0  
10:00 201.8 F 0.0  
15:00 209.1 F 0.0  
20:00 217.2 F 0.0  
25:00 217.2 F 0.0  
30:00 212.1 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 30:43  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:08:40  
CYCLE COMPLETE

02/06/2023 09:47 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_