

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1505

Name: RITTER2  
Sterilizer ID: V2313228  
09/20/2022 08:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.7 F 0.0  
2:00 161.7 F 1.7  
4:00 165.2 F 5.2  
6:00 193.3 F 10.3  
8:00 211.4 F 5.2  
10:00 227.8 F 10.1  
12:00 233.0 F 8.3  
14:00 241.9 F 12.9  
16:00 241.5 F 11.6  
18:00 254.3 F 18.2  
20:00 266.0 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.6 F 29.7  
2:00 271.6 F 29.0  
3:00 272.7 F 29.9  
4:00 272.1 F 29.2  
  
Min 271.5 F 29.0  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.5 F 0.0  
5:00 210.8 F 0.0  
10:00 219.2 F 0.0  
15:00 218.7 F 0.0  
20:00 218.6 F 0.0  
25:00 220.4 F 0.0  
30:00 191.0 F 0.0  
Drying Complete

FILLING: 1:43  
HEATING: 21:38  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:00:29  
CYCLE COMPLETE

09/20/2022 09:02 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_