

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4241

Name: RITTER 1  
Sterilizer ID: V2178198  
01/29/2024 12:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.9 F 0.0  
2:00 172.4 F 3.2  
4:00 183.2 F 7.3  
6:00 206.2 F 12.6  
8:00 221.8 F 6.6  
10:00 235.0 F 11.6  
12:00 234.9 F 8.5  
14:00 241.8 F 11.0  
16:00 245.4 F 12.9  
18:00 257.7 F 19.3  
20:00 268.4 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.0  
1:00 272.5 F 28.6  
2:00 271.2 F 28.0  
3:00 272.4 F 28.7  
4:00 271.5 F 28.0

Min 271.1 F 27.9  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.1 F 0.0  
5:00 224.1 F 0.0  
10:00 211.5 F 0.0  
15:00 207.9 F 0.0  
20:00 204.3 F 0.0  
25:00 214.5 F 0.0  
30:00 235.5 F 0.0  
Drying Complete

FILLING: 1:26  
HEATING: 21:13  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:00:03  
CYCLE COMPLETE

01/29/2024 01:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_