
Midmark M11 - v1.0.0
Total Cycles: 3421

Name: RITTER 1
Sterilizer ID: V2178198
06/21/2023 10:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.1 F 0.0
2:00 183.6 F 2.7
4:00 197.8 F 6.6
6:00 213.0 F 12.1
8:00 224.4 F 6.6
10:00 238.6 F 11.9
12:00 236.0 F 9.0
14:00 239.9 F 9.7
16:00 247.6 F 13.9
18:00 259.6 F 20.4
20:00 270.3 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.4 F 28.7
2:00 271.4 F 28.1
3:00 272.5 F 28.8
4:00 271.6 F 28.1

Min 271.4 F 28.0
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 235.6 F 0.0
10:00 229.7 F 0.0
15:00 228.5 F 0.0
20:00 229.2 F 0.0
25:00 227.8 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 20:53
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:38
CYCLE COMPLETE

06/21/2023 11:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____