

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 006735  
11:20 13/12/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:49  
PRESSURIZING 2:52  
STERILIZING 3:53  
136.4 C 222kPa 3:53  
136.4 C 224kPa 4:23  
136.4 C 220kPa 4:53  
136.4 C 224kPa 5:23  
136.3 C 222kPa 5:53  
136.3 C 222kPa 6:23  
136.4 C 223kPa 6:53  
136.3 C 222kPa 7:23  
136.3 C 222kPa 7:23  
VENTING 7:23  
AIR DRYING 8:14  
CYCLE COMPLETE 8:19

-----  
Digital Signature #  
94B8CC417CF0CE1F  
-----