
Midmark M11 - v1.0.5
Total Cycles: 4081

Name: RITTER2
Sterilizer ID: V2313228
11/04/2024 02:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.2 F 0.0
2:00 123.7 F 0.0
4:00 146.1 F 1.6
6:00 172.3 F 5.0
8:00 191.6 F 8.9
10:00 208.1 F 13.1
12:00 219.7 F 6.7
14:00 231.2 F 11.2
16:00 231.2 F 7.9
18:00 243.4 F 13.4
20:00 242.8 F 12.1
22:00 254.5 F 18.2
24:00 265.1 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.6 F 29.7
2:00 271.7 F 29.1
3:00 271.7 F 29.5
4:00 272.5 F 29.6

Min 271.2 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.8 F 0.0
5:00 219.5 F 0.0
10:00 232.4 F 0.0
15:00 220.1 F 0.0
20:00 222.8 F 0.0
25:00 235.7 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:59
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:04:38
CYCLE COMPLETE

11/04/2024 03:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____