
Midmark M11 - v1.0.0
Total Cycles: 1342

Name: RITTER 1
Sterilizer ID: V2178198
11/08/2021 03:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.1 F 0.0
2:00 194.9 F 2.9
4:00 212.7 F 7.6
6:00 226.8 F 12.9
8:00 229.7 F 8.1
10:00 240.9 F 13.2
12:00 239.7 F 11.0
14:00 238.2 F 9.5
16:00 251.8 F 16.2
18:00 263.5 F 23.0
20:00 271.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.3
1:00 272.4 F 28.7
2:00 272.3 F 28.9
3:00 272.0 F 28.4
4:00 272.7 F 29.1

Min 271.6 F 28.2
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.2 F 0.0
5:00 215.6 F 0.0
10:00 221.7 F 0.0
15:00 218.3 F 0.0
20:00 219.6 F 0.0
25:00 219.1 F 0.0
30:00 218.5 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 20:04
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

11/08/2021 04:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____