
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002429
10:54 17/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:09
PRESSURIZING 2:17
STERILIZING 3:22
136.4 C 224kPa 3:22
136.4 C 226kPa 3:52
136.4 C 225kPa 4:22
136.4 C 225kPa 4:52
136.1 C 224kPa 5:22
136.4 C 226kPa 5:52
136.4 C 223kPa 6:22
136.2 C 223kPa 6:52
136.2 C 223kPa 6:52
VENTING 6:52
AIR DRYING 7:45
CYCLE COMPLETE 36:15

Digital Signature #
5CB18A79731331FC
