

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2849

Name: RITTER 1  
Sterilizer ID: V2178198  
02/02/2023 12:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 206.8 F 0.0  
2:00 199.7 F 4.3  
4:00 209.1 F 9.4  
6:00 216.5 F 6.9  
8:00 230.5 F 9.9  
10:00 232.7 F 8.3  
12:00 244.2 F 13.4  
14:00 243.8 F 12.4  
16:00 257.1 F 19.4  
18:00 268.8 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 272.7 F 29.2  
2:00 271.7 F 28.5  
3:00 272.7 F 29.3  
4:00 272.0 F 28.7  
  
Min 271.6 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.5 F 0.0  
5:00 233.0 F 0.0  
10:00 234.7 F 0.0  
15:00 232.7 F 0.0  
20:00 226.6 F 0.0  
25:00 226.5 F 0.0  
30:00 227.8 F 0.0  
  
Drying Complete

FILLING: 1:46  
HEATING: 19:06  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:07  
CYCLE COMPLETE

02/02/2023 01:42 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_