
Midmark M11 - v1.0.0
Total Cycles: 3328

Name: RITTER 1
Sterilizer ID: V2178198
06/02/2023 12:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.0 F 0.0
2:00 132.1 F 0.0
4:00 149.1 F 0.0
6:00 180.1 F 3.2
8:00 205.4 F 6.7
10:00 218.2 F 10.7
12:00 222.2 F 5.9
14:00 231.4 F 9.7
16:00 236.8 F 9.8
18:00 241.1 F 11.7
20:00 239.2 F 10.0
22:00 251.5 F 16.1
24:00 262.3 F 22.3
26:00 272.4 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.5 F 28.9
2:00 271.1 F 28.2
3:00 272.6 F 29.0
4:00 271.7 F 28.5

Min 271.0 F 28.2
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 205.6 F 0.0
10:00 216.8 F 0.0
15:00 226.5 F 0.0
20:00 220.1 F 0.0
25:00 224.3 F 0.0
30:00 224.8 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 26:27
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:04:57
CYCLE COMPLETE

06/02/2023 01:55 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____