

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3690

Name: RITTER2  
Sterilizer ID: V2313228  
06/12/2024 01:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.2 F 0.0  
2:00 159.3 F 2.6  
4:00 175.5 F 6.2  
6:00 195.9 F 10.5  
8:00 226.8 F 6.3  
10:00 230.5 F 9.2  
12:00 237.0 F 10.6  
14:00 240.7 F 11.7  
16:00 239.3 F 10.3  
18:00 251.5 F 16.5  
20:00 262.6 F 22.9  
22:00 272.5 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.6  
1:00 272.0 F 29.6  
2:00 272.4 F 29.9  
3:00 272.5 F 30.1  
4:00 271.5 F 29.9

Min 271.3 F 29.5  
Max 273.1 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.7 F 0.0  
5:00 236.0 F 0.0  
10:00 230.6 F 0.0  
15:00 211.2 F 0.0  
20:00 224.8 F 0.0  
25:00 177.5 F 0.0  
30:00 170.8 F 0.0  
Drying Complete

FILLING: 1:24  
HEATING: 23:08  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:02:14  
CYCLE COMPLETE

06/12/2024 02:10 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_