
Midmark M11 - v1.0.5
Total Cycles: 195

Name: RITTER2
Sterilizer ID: V2313228
07/12/2021 12:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.6 F 0.0
2:00 202.8 F 4.2
4:00 207.7 F 10.2
6:00 222.8 F 6.8
8:00 235.1 F 12.2
10:00 237.7 F 10.5
12:00 237.3 F 9.5
14:00 251.6 F 16.5
16:00 263.2 F 23.3
18:00 272.2 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 271.9 F 29.0
2:00 272.3 F 29.3
3:00 272.5 F 29.7
4:00 271.6 F 29.1

Min 271.6 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 222.8 F 0.0
10:00 228.1 F 0.0
15:00 229.8 F 0.0
20:00 228.4 F 0.0
25:00 227.4 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 2:30
HEATING: 18:06
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:46
CYCLE COMPLETE

07/12/2021 01:52 PM
ACCEPTED: X
REJECTED:
Operator: U01