

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1946

Name: RITTER 1  
Sterilizer ID: V2178198  
05/18/2022 01:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.2 F 0.0  
2:00 175.8 F 4.1  
4:00 185.9 F 8.5  
6:00 214.6 F 11.9  
8:00 223.2 F 8.2  
10:00 238.0 F 12.9  
12:00 237.8 F 10.5  
14:00 236.1 F 8.4  
16:00 250.1 F 15.4  
18:00 262.1 F 22.2  
20:00 272.8 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.4 F 29.0  
2:00 271.5 F 28.7  
3:00 272.4 F 29.0  
4:00 271.6 F 28.7  
  
Min 271.4 F 28.4  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.2 F 0.0  
5:00 217.2 F 0.0  
10:00 232.6 F 0.0  
15:00 231.5 F 0.0  
20:00 232.1 F 0.0  
25:00 223.6 F 0.0  
30:00 218.6 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 20:22  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:50  
CYCLE COMPLETE

05/18/2022 02:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_