
Midmark M11 - v1.0.5
Total Cycles: 2684

Name: RITTER2
Sterilizer ID: V2313228
07/24/2023 08:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.7 F 0.0
2:00 74.1 F 0.0
4:00 111.6 F 0.0
6:00 106.2 F 0.0
8:00 122.1 F 0.2
10:00 156.1 F 0.7
12:00 176.9 F 2.9
14:00 190.5 F 5.1
16:00 199.7 F 7.9
18:00 208.2 F 11.2
20:00 217.3 F 6.6
22:00 223.9 F 8.9
24:00 233.6 F 12.8
26:00 235.9 F 9.9
28:00 239.5 F 10.2
30:00 245.7 F 13.3
32:00 255.4 F 18.5
34:00 264.5 F 24.0
36:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.8
1:00 272.1 F 29.7
2:00 272.4 F 29.7
3:00 271.5 F 29.2
4:00 272.3 F 30.0

Min 270.9 F 28.6
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.8 F 0.0
5:00 231.9 F 0.0
10:00 211.8 F 0.0
15:00 218.6 F 0.0
20:00 222.6 F 0.0
25:00 221.5 F 0.0
30:00 220.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 36:21
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:14:56
CYCLE COMPLETE

07/24/2023 09:30 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____