

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000927  
13:28 11/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:54  
PRESSURIZING 3:07  
STERILIZING 4:37  
136.4 C 230kPa 4:37  
136.3 C 232kPa 5:07  
136.2 C 231kPa 5:37  
136.4 C 232kPa 6:07  
136.4 C 231kPa 6:37  
136.1 C 228kPa 7:07  
136.5 C 233kPa 7:37  
136.4 C 232kPa 8:07  
136.4 C 232kPa 8:07  
VENTING 8:07  
AIR DRYING 9:01  
CYCLE COMPLETE 24:01

-----  
Digital Signature #  
4014DB32FF750539  
-----