
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003485
7:58 20/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:18
PRESSURIZING 1:24
STERILIZING 2:21
136.4 C 221kPa 2:21
136.4 C 223kPa 2:51
136.4 C 223kPa 3:21
136.2 C 222kPa 3:51
136.4 C 218kPa 4:21
136.2 C 217kPa 4:51
136.2 C 217kPa 5:21
136.3 C 222kPa 5:51
136.3 C 222kPa 5:51
VENTING 5:51
AIR DRYING 6:44
CYCLE COMPLETE 6:46

Digital Signature #
0987E39928AAF53A
