
Midmark M11 - v1.0.0
Total Cycles: 5244

Name: RITTER 1
Sterilizer ID: V2178198
01/15/2025 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 73.8 F 0.0
4:00 95.5 F 0.0
6:00 126.3 F 0.0
8:00 155.4 F 0.1
10:00 182.1 F 0.3
12:00 203.1 F 0.6
14:00 217.9 F 3.4
16:00 225.7 F 6.1
18:00 233.2 F 9.1
20:00 241.2 F 12.9
22:00 237.8 F 9.5
24:00 239.7 F 9.7
26:00 246.9 F 13.3
28:00 244.2 F 12.0
30:00 256.8 F 18.5
32:00 267.8 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.9
1:00 272.8 F 28.6
2:00 271.8 F 27.8
3:00 272.5 F 28.5
4:00 272.6 F 28.4

Min 271.6 F 27.7
Max 272.9 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.7 F 0.0
5:00 212.2 F 0.0
10:00 196.7 F 0.0
15:00 201.2 F 0.0
20:00 218.6 F 0.0
25:00 234.3 F 0.0
30:00 224.1 F 0.0

Drying Complete

FILLING: 0:43
HEATING: 33:23
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:11:34
CYCLE COMPLETE

01/15/2025 10:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____