

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1066

Name: RITTER 1  
Sterilizer ID: V2178198  
07/28/2021 11:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.1 F 0.0  
2:00 185.9 F 3.0  
4:00 212.9 F 7.9  
6:00 231.3 F 12.1  
8:00 230.4 F 8.8  
10:00 239.9 F 10.6  
12:00 242.0 F 11.8  
14:00 240.0 F 10.3  
16:00 253.4 F 17.0  
18:00 264.8 F 23.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.6 F 28.9  
2:00 271.3 F 28.2  
3:00 272.5 F 29.0  
4:00 271.8 F 28.4  
  
Min 271.2 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.6 F 0.0  
5:00 225.7 F 0.0  
10:00 199.4 F 0.0  
15:00 194.3 F 0.0  
20:00 190.0 F 0.0  
25:00 187.2 F 0.0  
30:00 186.0 F 0.0  
Drying Complete

FILLING: 1:56  
HEATING: 19:53  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:59  
CYCLE COMPLETE

07/28/2021 12:40 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01