
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003916
12:18 29/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:20
PRESSURIZING 1:23
STERILIZING 2:21
136.0 C 220kPa 2:21
136.4 C 224kPa 2:51
136.4 C 224kPa 3:21
136.3 C 223kPa 3:51
136.2 C 223kPa 4:21
136.4 C 224kPa 4:51
136.2 C 224kPa 5:21
136.4 C 224kPa 5:51
136.4 C 224kPa 5:51
VENTING 5:51
AIR DRYING 7:18
CYCLE COMPLETE 7:46

Digital Signature #
F38674771A6A020A
