
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001007
11:40 16/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:16
PRESSURIZING 2:24
STERILIZING 3:23
136.0 C 228kPa 3:23
136.3 C 232kPa 3:53
136.2 C 231kPa 4:23
136.1 C 230kPa 4:53
136.3 C 231kPa 5:23
136.3 C 232kPa 5:53
136.4 C 233kPa 6:23
136.4 C 233kPa 6:53
136.4 C 233kPa 6:53
VENTING 6:53
AIR DRYING 7:44
CYCLE COMPLETE 8:25

Digital Signature #
845F07771E8A6E80
