
Midmark M11 - v1.0.0
Total Cycles: 706

Name: MIDMARK
Sterilizer ID: V2178198
02/26/2021 07:44 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.1 F 0.0
2:00 82.3 F 0.0
4:00 80.6 F 0.0
6:00 109.6 F 0.0
8:00 143.6 F 0.2
10:00 177.1 F 0.3
12:00 210.6 F 2.5
14:00 215.1 F 5.6
16:00 222.8 F 8.5
18:00 232.2 F 12.1
20:00 231.3 F 7.6
22:00 241.9 F 12.7
24:00 240.2 F 10.4
26:00 236.5 F 8.5
28:00 250.7 F 15.2
30:00 261.9 F 21.6
32:00 272.4 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 27.9
1:00 272.6 F 28.5
2:00 271.3 F 27.8
3:00 272.6 F 28.6

Min 271.3 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.8 F 0.0
5:00 218.7 F 0.0
10:00 221.6 F 0.0
15:00 192.1 F 0.0
20:00 217.3 F 0.0
25:00 228.8 F 0.0
30:00 210.9 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 32:28
STERILIZING: 3:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:37
CYCLE COMPLETE

02/26/2021 08:53 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____