
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 001436
14:14 09/15/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:15
PRESSURIZING 7:10
STERILIZING 8:47
136.0 C 220kPa 8:47
136.2 C 221kPa 9:17
136.1 C 221kPa 9:47
136.3 C 222kPa 10:17
136.5 C 223kPa 10:47
136.0 C 221kPa 11:17
135.7 C 218kPa 11:47
136.4 C 223kPa 12:17
VENTING 12:17
AIR DRYING 13:23
CYCLE COMPLETE 43:23

Digital Signature #
8689096999D5A0A7
