

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2646

Name: RITTER2  
Sterilizer ID: V2313228  
07/14/2023 10:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 152.0 F 0.0  
2:00 154.6 F 0.0  
4:00 157.5 F 2.3  
6:00 180.4 F 6.3  
8:00 200.9 F 10.6  
10:00 215.7 F 5.6  
12:00 226.1 F 8.6  
14:00 237.0 F 13.1  
16:00 237.0 F 10.2  
18:00 236.8 F 9.1  
20:00 247.6 F 14.4  
22:00 258.5 F 20.5  
24:00 268.4 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.7 F 29.6  
2:00 271.6 F 28.9  
3:00 272.5 F 29.8  
4:00 272.2 F 29.3

Min 271.4 F 28.9  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.3 F 0.0  
5:00 205.9 F 0.0  
10:00 209.9 F 0.0  
15:00 210.3 F 0.0  
20:00 208.4 F 0.0  
25:00 208.1 F 0.0  
30:00 209.7 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 25:20  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:03:36  
CYCLE COMPLETE

07/14/2023 11:12 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_