

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.9uS / 0.5ppm  
CYCLE NUMBER 001718  
10:54 09/06/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:24  
PRESSURIZING 2:29  
STERILIZING 3:28  
136.4 C 220kPa 3:28  
136.4 C 224kPa 3:58  
136.3 C 223kPa 4:28  
136.4 C 223kPa 4:58  
136.4 C 224kPa 5:28  
136.3 C 223kPa 5:58  
136.3 C 223kPa 6:28  
136.4 C 225kPa 6:58  
136.4 C 225kPa 6:58  
VENTING 6:58  
AIR DRYING 7:52  
CYCLE COMPLETE 7:58

-----  
Digital Signature #  
D1C9211827EFAF36  
-----