
Midmark M11 - v1.0.0
Total Cycles: 4428

Name: RITTER 1
Sterilizer ID: V2178198
04/05/2024 08:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.1 F 0.0
2:00 91.0 F 0.3
4:00 129.0 F 1.2
6:00 111.8 F 3.4
8:00 148.2 F 6.3
10:00 173.9 F 9.6
12:00 196.9 F 13.2
14:00 209.7 F 3.8
16:00 220.3 F 6.9
18:00 231.7 F 10.4
20:00 230.8 F 8.5
22:00 237.7 F 10.3
24:00 235.6 F 8.6
26:00 245.8 F 13.3
28:00 256.9 F 19.2
30:00 267.3 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.4
1:00 272.7 F 29.1
2:00 271.1 F 28.3
3:00 272.7 F 29.2
4:00 271.7 F 28.5

Min 271.0 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.5 F 0.0
5:00 227.8 F 0.0
10:00 205.9 F 0.0
15:00 206.2 F 0.0
20:00 206.8 F 0.0
25:00 210.9 F 0.0
30:00 210.9 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:31
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:09:45
CYCLE COMPLETE

04/05/2024 09:24 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____