

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2380

Name: RITTER2  
Sterilizer ID: V2313228  
05/16/2023 04:53 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.8 F 0.0  
2:00 188.5 F 3.9  
4:00 203.9 F 9.2  
6:00 222.2 F 7.8  
8:00 232.7 F 11.2  
10:00 234.9 F 9.4  
12:00 237.1 F 9.3  
14:00 249.9 F 15.8  
16:00 262.5 F 23.1  
18:00 272.1 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.3 F 29.6  
2:00 271.3 F 29.0  
3:00 272.4 F 29.8  
4:00 271.8 F 29.2  
  
Min 271.3 F 28.9  
Max 272.5 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.5 F 0.0  
5:00 220.5 F 0.0  
10:00 234.3 F 0.0  
15:00 233.7 F 0.0  
20:00 228.1 F 0.0  
25:00 228.5 F 0.0  
30:00 226.2 F 0.0  
Drying Complete

FILLING: 2:47  
HEATING: 18:10  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:15  
CYCLE COMPLETE

05/16/2023 05:51 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_