
Midmark M11 - v1.0.5
Total Cycles: 2737

Name: RITTER2
Sterilizer ID: V2313228
08/04/2023 06:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.1 F 0.0
2:00 134.0 F 0.0
4:00 156.1 F 2.3
6:00 185.4 F 6.3
8:00 203.8 F 10.6
10:00 217.2 F 5.6
12:00 227.6 F 9.1
14:00 235.9 F 11.2
16:00 240.4 F 11.8
18:00 239.9 F 10.7
20:00 253.0 F 17.3
22:00 264.5 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.2 F 29.1
2:00 272.7 F 29.6
3:00 271.4 F 29.1
4:00 272.5 F 29.3

Min 271.3 F 28.9
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.9 F 0.0
5:00 203.2 F 0.0
10:00 196.3 F 0.0
15:00 194.8 F 0.0
20:00 193.7 F 0.0
25:00 195.3 F 0.0
30:00 202.7 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:53
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:02:18
CYCLE COMPLETE

08/04/2023 07:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____