
Midmark M11 - v1.0.5
Total Cycles: 2742

Name: RITTER2
Sterilizer ID: V2313228
08/09/2023 10:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.2 F 0.0
2:00 167.5 F 0.0
4:00 187.7 F 2.1
6:00 210.8 F 5.9
8:00 224.8 F 10.6
10:00 224.9 F 5.9
12:00 235.3 F 10.0
14:00 236.6 F 9.1
16:00 242.1 F 12.2
18:00 239.5 F 10.5
20:00 251.4 F 16.6
22:00 262.0 F 22.8
24:00 271.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.2 F 30.0
2:00 272.6 F 29.8
3:00 271.6 F 29.2
4:00 271.3 F 29.4

Min 271.2 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 213.3 F 0.0
10:00 227.6 F 0.0
15:00 219.1 F 0.0
20:00 218.5 F 0.0
25:00 228.7 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 24:39
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:02:44
CYCLE COMPLETE

08/09/2023 11:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____