

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001606  
16:43 16/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:04  
PRESSURIZING 2:08  
STERILIZING 3:09  
136.5 C 225kPa 3:09  
136.3 C 226kPa 3:39  
136.4 C 224kPa 4:09  
136.3 C 226kPa 4:39  
136.3 C 224kPa 5:09  
136.4 C 226kPa 5:39  
136.4 C 224kPa 6:09  
136.4 C 227kPa 6:39  
136.4 C 227kPa 6:39  
VENTING 6:39  
AIR DRYING 7:31  
CYCLE COMPLETE 7:40  
-----

Digital Signature #  
E421A42C1E0A544F  
-----