

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4432

Name: RITTER2  
Sterilizer ID: V2313228  
03/17/2025 11:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 82.0 F 0.0  
2:00 83.1 F 0.3  
4:00 92.4 F 1.3  
6:00 119.2 F 3.5  
8:00 149.5 F 6.3  
10:00 174.4 F 9.2  
12:00 192.2 F 12.5  
14:00 212.9 F 4.8  
16:00 221.8 F 8.5  
18:00 232.8 F 12.5  
20:00 230.7 F 7.9  
22:00 241.1 F 12.8  
24:00 239.7 F 10.6  
26:00 250.5 F 16.1  
28:00 260.3 F 21.7  
30:00 269.8 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 272.1 F 30.0  
2:00 271.7 F 29.5  
3:00 272.5 F 29.7  
4:00 272.1 F 30.1  
  
Min 271.0 F 29.0  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.5 F 0.0  
5:00 178.9 F 0.0  
10:00 190.4 F 0.0  
15:00 195.7 F 0.0  
20:00 191.6 F 0.0  
25:00 196.7 F 0.0  
30:00 207.8 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 30:55  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:09:41  
CYCLE COMPLETE

03/17/2025 01:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_