
Midmark M11 - v1.0.5
Total Cycles: 1994

Name: RITTER2
Sterilizer ID: V2313228
02/01/2023 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.8 F 0.0
2:00 138.9 F 0.1
4:00 170.1 F 3.4
6:00 186.3 F 7.9
8:00 206.2 F 13.0
10:00 218.2 F 8.0
12:00 233.3 F 13.3
14:00 236.7 F 10.7
16:00 236.0 F 9.2
18:00 250.7 F 16.1
20:00 262.6 F 23.1
22:00 272.6 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.4 F 29.9
2:00 271.6 F 29.1
3:00 271.5 F 29.7
4:00 272.4 F 29.7

Min 271.1 F 28.9
Max 272.4 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.5 F 0.0
5:00 198.2 F 0.0
10:00 219.2 F 0.0
15:00 225.3 F 0.0
20:00 229.0 F 0.0
25:00 233.5 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 22:15
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:25
CYCLE COMPLETE

02/01/2023 02:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____