
Midmark M11 - v1.0.0
Total Cycles: 3270

Name: RITTER 1
Sterilizer ID: V2178198
05/23/2023 02:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.8 F 0.0
2:00 190.1 F 0.9
4:00 215.9 F 6.3
6:00 224.0 F 11.8
8:00 230.0 F 7.7
10:00 242.4 F 13.4
12:00 241.7 F 11.4
14:00 240.0 F 10.2
16:00 253.9 F 17.1
18:00 265.5 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.5 F 28.7
2:00 271.5 F 28.0
3:00 272.6 F 28.9
4:00 272.0 F 28.3

Min 271.5 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 230.0 F 0.0
10:00 232.2 F 0.0
15:00 227.4 F 0.0
20:00 226.0 F 0.0
25:00 226.2 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 19:44
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

05/23/2023 03:32 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____