

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5113

Name: RITTER 1  
Sterilizer ID: V2178198  
11/20/2024 11:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.4 F 0.0  
2:00 175.1 F 0.0  
4:00 182.8 F 0.5  
6:00 211.1 F 1.7  
8:00 220.1 F 6.0  
10:00 227.2 F 10.4  
12:00 232.4 F 7.9  
14:00 242.2 F 12.3  
16:00 240.0 F 10.5  
18:00 237.3 F 9.1  
20:00 251.4 F 15.8  
22:00 262.6 F 22.2  
24:00 272.8 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.1  
1:00 272.4 F 28.9  
2:00 272.2 F 28.5  
3:00 271.5 F 28.0  
4:00 271.6 F 28.5

Min 271.3 F 28.0  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.3 F 0.0  
5:00 212.4 F 0.0  
10:00 228.1 F 0.0  
15:00 235.5 F 0.0  
20:00 230.6 F 0.0  
25:00 224.3 F 0.0  
30:00 224.5 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 24:27  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:03:08  
CYCLE COMPLETE

11/20/2024 12:16 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_