
Midmark M11 - v1.0.5
Total Cycles: 2412

Name: RITTER2
Sterilizer ID: V2313228
05/25/2023 07:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.4 F 0.0
2:00 87.9 F 0.0
4:00 88.7 F 0.0
6:00 109.8 F 0.0
8:00 160.6 F 1.1
10:00 177.3 F 4.7
12:00 200.0 F 8.5
14:00 220.3 F 12.4
16:00 223.7 F 6.2
18:00 233.8 F 10.6
20:00 233.4 F 8.0
22:00 243.9 F 13.0
24:00 242.5 F 11.6
26:00 254.7 F 18.0
28:00 265.8 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.5 F 29.2
2:00 271.2 F 28.6
3:00 272.6 F 29.3
4:00 271.4 F 28.5

Min 271.0 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.9 F 0.0
5:00 208.6 F 0.0
10:00 217.1 F 0.0
15:00 221.4 F 0.0
20:00 220.7 F 0.0
25:00 222.6 F 0.0
30:00 223.6 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 29:44
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:07:41
CYCLE COMPLETE

05/25/2023 08:46 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____