
Midmark M11 - v1.0.5
Total Cycles: 1982

Name: RITTER2
Sterilizer ID: V2313228
01/30/2023 05:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.0 F 0.0
2:00 176.2 F 4.0
4:00 193.9 F 8.9
6:00 212.5 F 7.7
8:00 227.0 F 9.6
10:00 233.0 F 8.7
12:00 243.6 F 13.5
14:00 244.2 F 12.8
16:00 257.4 F 19.9
18:00 269.1 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.5 F 29.6
2:00 271.5 F 29.2
3:00 272.5 F 29.7
4:00 271.7 F 29.0

Min 271.5 F 29.0
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 224.1 F 0.0
10:00 231.9 F 0.0
15:00 231.9 F 0.0
20:00 225.0 F 0.0
25:00 226.4 F 0.0
30:00 231.5 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 19:02
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:07
CYCLE COMPLETE

01/30/2023 06:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____